

## V-Band Spectrum Analyzer Mixer, Common LO/IF Port

### Description:

**Model STH-15SF-S1** is a V Band harmonic mixer employs single diode and broadband circuitry to deliver low conversion loss and continuous frequency coverage for full waveguide band operation covering the frequency range of 50 to 75 GHz. The mixer extends the operation frequency of the test instruments like the spectrum analyzer and frequency counter from 18 GHz or below to full V band operations. Unlike the balanced harmonic mixer, model number SFH-15SFSF-A3, the spectrum analyzer mixer does not have a built-in frequency diplexer. Therefore, their LO and IF ports are combined to a common coaxial port where LO and IF signals share. This feature provides a convenient connection ability when used with the spectrum analyzer or frequency counter models having built-in diplexer, i.e., combined LO and IF port, such as the models offered by Advantest, Anritsu, Rohde & Schwarz, some Keysight models and Tektronix and National Instruments/Phase Matrix (EIP).



### Features:

- Full Waveguide Coverage
- Combined LO and IF Port
- Instrumentation Grade

### Applications:

- Phase Lock Loops
- Spectrum Analyzer with built-in Diplexer
- Frequency Counter with built-in Diplexer

### Electrical Specifications:


Parameter	Minimum	Typical	Maximum
Frequency	50 GHz		75 GHz
IF Frequency	DC	1.0 GHz	4.0 GHz
LO Power	+10 dBm	+13 dBm	+16 dBm
Sensitivity		-85 dBm @ 1 KHz RBW	
RF Input Power		-20 dBm	+16 dBm
Specification Temperature		+25 °C	
Operating Temperature	0 °C		+50 °C

### Mechanical Specifications:

Item	Specification
RF Port	WR-15 Waveguide with UG-385/U Flange
IF and LO Ports	SMA (F)
Material	Aluminum
Finishing	Gold Plated
Weight	1 Oz
Outline	FD-V1

## V-Band Spectrum Analyzer Mixer, Common LO/IF Port

**Mechanical Outline:** (Unless otherwise specified, all dimensions are in inches)



**Note:**

- SAGE Millimeter, Inc. reserves the right to change the information presented without notice.

**Caution:**

- Exceeding absolute maximum ratings of the mixer will damage the detector.
- The mixer is a static sensitive device. Always follow ESD rules when working with the mixer.
- Any foreign objects into the waveguide will lower performance and/or damage device.
- Proper torque,  $8.0 \pm 0.15$  inch-pounds ( $0.92 \pm 0.05$  Nm), should be applied. **SAGE Millimeter torque wrench, model SCH-08008-S1, is highly recommended.**